

Memo on HODF24

- HODF24
 - covering whole exit area of FDC2
 - 8 scintillator bars added to HODF(16)
 - budget from Tohoku Univ.

- HODF to HODF24
 - 8 scintillator slats were added to HODF
 - 24 units were installed on a new detector platform (25-Mar-2015 ~ 29-Mar-2015).
 - It can cover the exit window of FDC2 (~2300mm horizontally).
 - During the re-installation, 5 PMT's optical contacts were found to be broken. They are HODF-11U, 12U, 14U, 15U, and 16U, and are being fixed.
 - They will be installed into HODF24-1 to 5.
 - budget from Tohoku Univ.

HODP & HODF24



