

Memo for exp_samurai_2016fall

- works done on 28-Oct (Fri) @RIKEN
 - BDC1, 2, FDC1, FDC2
 - w & w/o test pulse into ASD : OK
 - Gas status
 - HV status

-
- BDC1, 2, FDC1, FDC2 : with Isobe-san
 - data taken with and without test pulse into ASD
 - with test input : no missing channels
 - without test input : no noise (accidental) hits
 - Gas/ vacuum
 - ICB
 - ~20 cc/min P10 since 21-Oct(Fri)
 - HV conditioning
 - @0.5 kV: 28-Oct~
 - FDC2
 - He/CH₄ : about 4.6 times exchange on 28-Oct(Fri)
 - HV conditioning
 - @0.4 kV : 21-Oct ~ 28-Oct
 - @1.0 kV : 28-Oct~
 - BDC1, BDC2, FDC1
 - low pressure : 22-Oct~
 - ICB
 - slight noise (probably from Turbo pump?)